HANAROSTENT ®
Duodenum/Pylorus Kim's Flare (NCN)



Indication :

• Palliative treatment of pyloric or duodenal stricture caused by malignant tumors

Features :

• Flare heads on the both ends for Anti-Migration
• Uncovered on the both ends for Anti-Migration
• Larger head and a lasso on the proximal
• Double layered body structure and high conformity

Fluoroscopic Application

	Stent (mm)			Delivery device	
	Diameter	Usable length	Total length*	Length(mm)	Diameter(mm/fr)
DPDL-20-060-180	22-20-40	24	60	1800	3.4/10.2
DPDL-20-150-180		114	150	1800	3.4/10.2

*Increment of 10mm

Fluoroscopic Application

	Stent (mm)			Delivery device	
	Diameter	Usable length	Total length*	Length(mm)	Diameter(mm/fr)
DPDL-20-060-120	22-20-40	24	60	1200	3.4/10.2
DPDL-20-150-120		114	150	1200	3.4/10.2

*Increment of 10mm

Distal  Proximal

(NNN) Non covered
(CCC) Fully covered
(NCN) Partially covered

XXX-00-000-000 (① ② ③ ④)
① Product Name
② Stent Diameter

XXX 00 000-X000 (① ② ③ ④)
③ Stent Length
④ Delivery Device Length

HANAROSTENT ®

Duodenum/Pylorus Kim's Flare (NCN)



Indication :

• Palliative treatment of pyloric or duodenal stricture caused by malignant tumors

Features :

• Flare heads on the both ends for Anti-Migration
• Uncovered on the both ends for Anti-Migration
• Larger head and a lasso on the proximal
• Double layered body structure and high conformity

Endoscopic Application

	Stent (mm)			Delivery device	
	Diameter	Usable length	Total length*	Length(mm)	Diameter(mm/fr)
DPDL-20-060-230	22-20-40	24	60	2300	3.4/10.2
DPDL-20-150-230		114	150	2300	3.4/10.2

*Increment of 10mm

Distal  Proximal

(NNN) Non covered
(CCC) Fully covered
(NCN) Partially covered

XXX-00-000-000
① ② ③ ④

① Product Name
② Stent Diameter

XXX 00 000-X000
① ② ③ ④

③ Stent Length
④ Delivery Device Length